



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/621,196	07/15/2003	Ychoshua Shachar	MNETEC.001A	2337
20995 7590 09/10/2007 KNOBBE MARTENS OLSON & BEAR LLP 2040 MAIN STREET FOURTEENTH FLOOR IRVINE, CA 92614			EXAMINER APANIUS, MICHAEL	
			ART UNIT 3736	PAPER NUMBER
			NOTIFICATION DATE 09/10/2007	DELIVERY MODE ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jcartee@kmob.com
eOAPilot@kmob.com

Office Action Summary	Application No. 10/621,196	Applicant(s) SHACHAR, YEHOASHUA	
	Examiner Michael Apanius	Art Unit 3736	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 15 June 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 39-43 and 45-47 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 39-43 and 45-47 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 15 June 2007 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>6/15/2007</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

1. The amendment to claim 39, the abstract, and the drawings are acknowledged. Reference 8 listed on the attached IDS was lined out because the reference is already of record in the current application.

Specification

2. The disclosure is objected to because of the following informality. The filing date of the provisional application stated in the first paragraph appears incorrect.

Appropriate correction is required.

3. The specification is objected to as failing to provide proper antecedent basis for the claimed subject matter. See 37 CFR 1.75(d)(1) and MPEP § 608.01(o). Correction of the following is required: the specification fails to provide proper antecedent basis for the connection via magnetic material as set forth in claim 43.

Claim Rejections - 35 USC § 112

4. The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

5. Claims 39-43 and 45-47 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably

Art Unit: 3736

convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. Amended claim 39 states, "said system controller compensates for a dynamic position of a wall of a heart chamber".

Although the original disclosure appears to support compensating for a dynamic position of a beating heart (i.e. see original abstract), the original disclosure does not appear to support compensating for a dynamic position of a *wall of a heart chamber*.

Therefore, the amendment does not appear to be properly supported by the original disclosure.

Claim Rejections - 35 USC § 103

6. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

7. Claims 39, 42 and 45 are rejected under 35 U.S.C. 103(a) as being unpatentable over Blume et al. (US 6,014,580) in view of Nowlin et al. (US 6,459,926) and Wang et al. (US 5,971,976).

8. Note that Blume uses the calculations of Werp et al. (US 6,015,414) and incorporates Werp by reference (see Blume, paragraph bridging columns 6 and 7).

9. Blume discloses an apparatus for controlling movement of a tool to be inserted into the body of a patient, comprising: a controllable magnetic field source (12) having a first cluster of electromagnet poles (X+, Y- in figure 4) and a second cluster of

Art Unit: 3736

electromagnet poles (X-, Y+), said first cluster of poles substantially opposed to said second cluster of poles; a tool (column 5, lines 18-21) having a distal end responsive to said magnetic field; one or more sensors (20) configured to sense a current position of said distal end; and a system controller (112) for controlling said magnetic field source to control a movement of said distal end according to a feedback calculation (see formulas in column 6 of Werp) wherein said system controller is configured to compute a position error comprising a difference (error/correction vectors of Werp) between a desired position of said distal end and said current position of said distal end. In regards to claim 42, the apparatus comprises an operator interface unit (column 8, lines 7-19).

10. Although Blume discloses a Virtual Tip control device (column 8, lines 7-19), Blume does not expressly disclose that the system controller computes an amount of tactile feedback according to the position error, nor does Blume expressly disclose a Virtual Tip Calibration Fixture.

11. Nowlin teaches computing an amount of tactile feedback according to a position error (column 19, lines 23-67) so that the master and slave controllers are in corresponding positions. Furthermore, the feedback will tactilely notify a surgeon that the surgical tool has encountered resistance to movement in a desired direction (column 19, last paragraph). In regards to claim 45, Nowlin further teaches a Virtual Tip Calibration Fixture (200).

12. Therefore, it would have been obvious to one having ordinary skill in the art at the time of invention to have modified the system controller of Blume to compute an

Art Unit: 3736

amount of tactile feedback according to the position error and to have used a calibration fixture as taught by Nowlin in order to ensure that the position of the Virtual Tip corresponds to the internal position of the tool and to notify a surgeon of resistance to movement in a particular direction.

13. Blume also does not expressly disclose a correction input to said desired position computed based on data from an auxiliary device, such that said system controller compensates for a dynamic position of a wall of a heart chamber such that said distal end moves substantially in unison with said wall.

14. Wang teaches a system controller (14) for a robotic surgical system (16, 18), wherein a correction input is computed based on data from an auxiliary device (52, 54), such that said system controller can compensate for a dynamic position of a wall of a heart chamber such that a distal end moves substantially in unison with said wall (column 6, line 40 - column 8, line 6). The correction input compensates and corrects for heart motion to enable surgeries to be carried out near a beating heart with minimal risk and operating time (column 2, lines 20-29).

15. Therefore, it would have been obvious to one having ordinary skill in the art at the time of invention to have used a correction input such that a tool distal end can move in unison with a wall of a heart chamber as taught by Wang in the apparatus of Blume as modified by Nowlin in order to allow surgeries to be performed near a moving heart with minimal risk and operating time by compensating and correcting for heart motion.

Art Unit: 3736

16. Claims 40 and 41 are rejected under 35 U.S.C. 103(a) as being unpatentable over Blume et al. (US 6,014,580) as modified by Nowlin et al. (US 6,459,926) and Wang et al. (US 5,971,976), as applied to claims 39, 42 and 45 above, and further in view of Solf et al. (US 6,587,709).

17. Blume as modified by Nowlin and Wang does not expressly disclose one or more piezoelectric rings.

18. Solf teaches using piezoelectric rings (10a, 10b, 10c) with an ultrasound emitter (1) for the purpose of automatically tracking a catheter tip without manual displacement of an ultrasound transducer (abstract; column 2, lines 39-46).

19. Therefore, it would have been obvious to one having ordinary skill in the art at the time of invention to have used piezoelectric rings as taught by Solf in the apparatus of Blume as modified by Nowlin and Wang in order to automatically and accurately track a catheter tip at all times.

20. Claim 43 is rejected under 35 U.S.C. 103(a) as being unpatentable over Blume et al. (US 6,014,580) as modified by Nowlin et al. (US 6,459,926) and Wang et al. (US 5,971,976), as applied to claims 39, 42 and 45 above, and further in view of Hastings (US 6,148,823).

21. Blume as modified by Nowlin and Wang does not expressly disclose that the first cluster of poles is connected to said second cluster of poles by a magnetic material.

22. Hastings teaches that magnetic poles can be connected by magnetic material for the purpose of providing a strong field for a given magnet cost (column 3, lines 39-42).

Art Unit: 3736

23. It would have been obvious to one having ordinary skill in the art at the time of invention to have connected the clusters of poles of Blume as modified by Nowlin and Wang with magnetic material as taught by Hastings in order to provide a stronger field for a given magnet.

24. Claims 46 and 47 are rejected under 35 U.S.C. 103(a) as being unpatentable over Blume et al. (US 6,014,580) as modified by Nowlin et al. (US 6,459,926) and Wang et al. (US 5,971,976), as applied to claims 39, 42 and 45 above, and further in view of Haynor et al. (US 6,129,668) and Tanabe et al. (US 5,550,469).

25. Blume further discloses a communication controller (54). However, Blume as modified by Nowlin and Wang does not expressly disclose one or more magnetic field sensors.

26. Haynor teaches one or more Hall-effect magnetic sensors to sense a position of a tool by sensing a magnetic field produced at the tool for the purpose of obviating the need to independently verify positioning with imaging equipment (column 2, lines 42-47).

27. It would have been obvious to one having ordinary skill in the art at the time of invention to have used one or more magnetic sensors as taught by Haynor in the apparatus of Blume as modified by Nowlin and Wang in order to determine the position of an indwelling tool without the need for imaging equipment.

28. Blume as modified by Nowlin, Wang and Haynor does not expressly disclose one or more temperature sensors.

Art Unit: 3736

29. Tanabe teaches a temperature-dependent variable resistor which is considered a temperature sensor for the purpose of compensating the temperature dependence of Hall-effect sensors (abstract). The temperature sensor is paired with magnetic sensors.

30. Therefore, it would have been obvious to one having ordinary skill in the art at the time of invention to have paired temperature sensors as taught by Tanabe with the Hall-effect sensors of Blume as modified by Nowlin, Wang and Haynor in order to compensate for the temperature dependence of Hall-effect sensors.

Response to Arguments

31. Applicant's arguments with respect to the previous prior art rejections have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

32. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Takasugi (US 4,735,211) discloses an ultrasonic measurement apparatus that compensates for heart movement.

33. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

Art Unit: 3736

34. A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

35. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Apanius whose telephone number is (571) 272-5537. The examiner can normally be reached on Mon-Fri 8am-4:30pm.

36. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on (571) 272-4726. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3736

37. Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MA

McHendry
MCHENDRY
PATENT EXAMINER
EBC CENTER 3700